

# VERIFICATION OF CALIBRATION REPORT

FEB 23 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130299</u>	
Name <u>Christopher M. Babcock</u> ID# <u>3867</u> Date <u>2-16-10</u>	
A	Agency <u>King Cove DPS</u> Phone # <u>497-2210</u>
Instrument Location <u>Squad Room</u>	
B	Alco S/N <u>X124634</u> Target Value <u>.078</u> High Pressure <u>1000 psi</u>
Alco Test Value Average <u>.077</u> <u>.077</u> 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>	
(OVER)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of March, 2010.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130299

FEBRUARY 16, 2010

OPERATOR'S NAME:  
BARCOCK CHRISTOPHER M  
OPERATOR'S NUMBER: 3867  
SUBJECT'S LAST NAME:

UOC  
SUBJECT'S FIRST NAME/MI :

UOC  
O.L. #: 123456  
DEPT/AGENCY: KUC1  
CASE/REPORT: UOC

TEST TYPE: U  
ALCO TARGET VALUE: .078 ✓  
ALCO S/N: X124634

### --- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.81 in  
ALCO TARGET .077 10:54  
BLANK TEST .000 10:55  
INTERNAL STANDARD VERIFIED 10:55  
ALCO TU 29.81 in .077 10:56  
BLANK TEST .000 10:57  
SUBJECT SAMPLE .000 10:57  
BLANK TEST .000 10:58  
ALCO TU 29.81 in .077 10:58  
BLANK TEST .000 10:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130299

FEBRUARY 16, 2010  
TIME 10:52

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c

BAROMETER: 29.81 in

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■